

8

Notice of Allowability

Application No.

10/027,939

Applicant(s)

DAI ET AL.

Examiner

Stefan Stojnov

Art Unit

2116

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to communication filed May 5, 2005.
2. ☒ The allowed claim(s) is/are 1-25.
3. ☒ The drawings filed on 20 December 2001 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

- | | |
|--|--|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892) | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 6. <input type="checkbox"/> Interview Summary (PTO-413),
Paper No./Mail Date _____. |
| 3. <input checked="" type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date <u>5/5/05 and 5/11/05</u> | 7. <input type="checkbox"/> Examiner's Amendment/Comment |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit
of Biological Material | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance |
| | 9. <input type="checkbox"/> Other _____. |

Allowable Subject Matter

Claims 1-25 are allowed.

The following is an examiner's statement of reasons for allowance:

Re claims 1, 12, 17, and 22:

Tatsu, Japanese Patent No. 11-353230, teaches flushing or maintaining of the cache contents upon entering a low power state (power-saving mode) depending on the residual electric energy of a dc-battery (paragraph 0017). However, in Tatsu there is no determination whether to flush or maintain the cache based on avoiding increased soft error rate as claimed.

Dai et al., U.S. Patent No. 6,347,379, teaches reducing leakage power consumption where in addition to lowering the voltage and stopping the clock, caches may be flushed to reduce the soft error rate (Abstract, FIG. 5). However, Dai does not teach maintaining of flushing the cache based on evaluating whether avoiding an increased soft error rate is a priority as claimed.

Fuller, U.S. Patent No. 5,632,038, teaches turning off and on the cache based on the processor's activity with additional logic determining if the cache needs to be flushed prior of turning of the cache (Abstract, FIGURE 3). There is no consideration for soft error rate verses lower power.

The prior art of record does not teach or suggest individually or in combination "a cache to flush or maintain its contents depending on a power status signal and an indication of whether a power reduction associated with maintaining the contents of the

Art Unit: 2116

cache upon entering the low power state is a lower priority than avoiding an increase in a soft error rate in the cache associated with reducing the voltage to the second level”.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled “Comments on Statement of Reasons for Allowance.”

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Stefan Stoykov whose telephone number is (571) 272-4236. The examiner can normally be reached between 8:00AM-4:30PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Lynne Browne can be reached on (571) 272-3670. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

SS


LYNNE H. BROWNE
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 2100